



**HAL**  
open science

## Organ-on-chip to investigate host-pathogens interactions

Thomas Feugas, Nathalie Sauvonnet

► **To cite this version:**

Thomas Feugas, Nathalie Sauvonnet. Organ-on-chip to investigate host-pathogens interactions. Cellular Microbiology, 2021, 23 (7), pp.e13336. 10.1111/cmi.13336 . pasteur-03677610

**HAL Id: pasteur-03677610**

**<https://pasteur.hal.science/pasteur-03677610>**

Submitted on 24 May 2022

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## **Organ-on-chip to investigate host-pathogens interactions**

Thomas Feaugas<sup>1</sup> and Nathalie Sauvonnet<sup>1\*</sup>

<sup>1</sup> Group Intracellular Trafficking and Tissue Homeostasis, Institut Pasteur, 28 rue du Dr Roux, 75015, Paris, France

\*Corresponding author Nathalie Sauvonnet

Email: [nathalie.sauvonnet@pasteur.fr](mailto:nathalie.sauvonnet@pasteur.fr)

### **Abstract**

Infectious diseases remain the subject of intense research. This topic reaches a new era towards the study of host-pathogen interactions mechanisms at the tissue scale. The past few years have hence witnessed the emergence of new methods. Among them, organ-on-chip, which combine biomaterial technology, microfluidic and tissue engineering to recreate the organ physiology is very promising. This review summarizes how this technology recapitulates the architecture, the mechanical stimulation and the interface of a tissue and how this particular microenvironment is critical to study host-pathogen interactions.

**Keywords** : Organ-on-chip, Microfluidic, Bacteria, Virus, Microbiota, Mechanical forces

### **1-Introduction**

Biomedical research aims at understanding cellular and molecular mechanisms of diseases to provide a suitable diagnosis or a possible cure. Methods to study deregulated human physiology have evolved and got widely improved in the last 50 years. Every tissue in the body behaves differently, has its own microenvironment, thereby complexifying a lot the methods to fully recapitulate their physiology. Even though studies performed on individual cells plated on classical petri dishes have enabled great discoveries, they are no longer sufficient to explore deeply complex mechanisms, forcing biomedical research to find new approaches.

When it comes to host-pathogen interactions, some concepts about pathogenic mechanisms still remain unclear and not completely explored yet. In particular, the role of mechanical forces

during infectious processes has been only recently investigated in few studies. Yet, these physical forces may modify the pathogen adhesion and virulence.

In that respect, our understanding of the role of the microenvironment and more specifically the biomechanical stimuli on host-pathogen interactions is still limited due to the lack of physiologically relevant organ models. The interface is considered as primordial since pathogen dynamics rely on mechanisms occurring at that precise location. The skin, the gastrointestinal tract, respiratory organs are examples of tissues which harbor interfaces with multiple microorganisms, some being pathogens. Therefore, deciphering host-pathogen interactions necessarily goes through the modeling of an accurate organ-like tissue as a working base. Tissue engineering has come in this context, as a huge breakthrough, offering broader perspectives with the design of organ-on-chip.

In this review, we will present different approaches to recapitulate the tissue physiology using organ-on-chip technology and how it can help to better investigate host-pathogen interactions at the tissue scale.

## **2-Organ-on-chip: concepts and designs**

Organ-on-chip emerged from the combination of microfluidic and tissue engineering. This recent state-of-the-art methodology does not aim to reconstitute the whole living organ but rather to synthesize the cell organization and physical parameters of an organ. The number of organ-on-chip models has grown exponentially and has been listed in several recent reviews (DOI: 10.1038/s41578-018-0034-7

DOI: 10.1126/science.aaw7894). There are many challenges associated to organ-on-chip technologies, depending to the system one study.

Organ-on-chip are microfluidic cell culture devices, engineered by microfabrication methods. The device made of clear and biocompatible polymers contains continuously perfused chambers, which can receive specific cell cultures. Today organ-on-chip microfluidic platforms are starting to reach the market and several companies such as Emulate™, Alevolix™, Mimetas™, and BiOND™ proposed such models. All chips have at least one channel providing various interface (air-liquid, liquid-liquid) and /or mechanical stimulation (shear, stretch) and recapitulating 3D or 2D tissue organization (examples are described in Figure 1). In most devices the microchannels are perfused with culture medium with flow rate leading to laminar fluid flows inside the channels. An advantage of these organ-on-chips is their ability to recover exit fluids and the

possibility to add any drugs or microorganisms in real time. Moreover, organ-on-chips are often made of transparent material enabling live imaging (DOI: 10.1016/j.chom.2019.08.007).

Several organ-on-chip designs differ from each other's by the material used, their ability to modulate physical forces or to recapitulate the 3D architecture of the organ.

## **2-1-Materials**

The choice of the material is highly important in the design process of organ-on-chip technologies. Biocompatible, non-toxic, optically clear for microscopic imaging are often preferred approaches. The stiffness of the material is decisive for cellular growth, one main challenge is to find the right balance between stiffer materials with more precise mechanical properties and softer materials (hydrogels) closer to physiological tissue properties, but less flexible for mechanical stimulation.

Polymers show many advantages since they can be shaped and designed with several techniques (photolithography, multiphoton excitation leading to local degradation, electrospinning...). Polydimethylsiloxane (PDMS) is the most common polymer used to build microfluidic organ-on-chip (Figure 1A, H. J. Kim & Ingber, 2013), classified as elastomer for its elasticity, with a high stiffness (800 kPa~10 MPa, (Seghir & Arscott, 2015)). In this context, a microengineered organ-on-chip incorporating a built-in perfusable vascular network has been described (Zhang et al., 2018) (Zhang & Radisic, 2017). This multi-material scaffolding (poly(octamethylene maleate (anhydride) citrate) and PDMS) named *AngioChip* is being cultivated with endothelial and parenchymal cells to model vascular systems (Figure 1D). Thanks to tunable mechanical properties and great biocompatibility, fabricated vascular networks demonstrate applications for human myocardium engineering.

Hydrogels are being investigated nowadays since they are highly biocompatible and many features like porosity, mechanical stiffness, or elasticity can be modulated (Ding et al., 2020). They are 3D polymers soft networks containing mostly water (Seliktar, 2012), with a lower but tunable stiffness (0.5 kPa ~ 100 kPa, (Zhao et al., 2019)) and can be classified into natural, synthetic and hybrid materials. Natural hydrogels exhibit closer to native tissues matrix but do not provide reproducible mechanical properties. Further considerations lead to finding the best choice between synthetic or hybrid hydrogels, guaranteeing a sufficient control on mechanical properties and a scaffold presenting interesting cell binding sites. As an example, two hydrogel based models mimicking intestinal crypts are shown Figure 1B (Nikolaev et al., 2020) and 1C(R. Kim et al., 2019). Some advanced techniques like stereolithographic high-resolution printing have enabled the development of such devices achieving precise localization and patterning of

cells. In light of their modular composition, hydrogels offer the possibility to create internal molecular gradients such as oxygen, which can be useful in particular for anaerobic microorganisms. In addition, hydrogels stiffness can be modulated in a reversible manner by crosslinking reactions following light stimulation (Rosales et al., 2017) thereby offering a remote spatio-temporal control interesting for mechanobiology issues.

### **2-2-Physical forces in microfluidic devices**

Organ and tissue inside the body experience physical changes, constraints within their own microenvironment and depending on their functions. Indeed, a blood vessel or the urinary tract will experience strong shear forces due to pulsatile blood and urinary flow rate but also some stretching forces during diastolic and systolic cycles in cardiac tissues. Intuitively, same forces can be considered in respiratory tissues (lung, pulmonary alveoli) since air is cyclically renewed during respiration. Some tissues such as the intestine are subjected to slower flows but experience similarly a cyclic stretch during peristalsis-like motions.

Moreover, mechanical forces stimulate cell differentiation and morphological polarization. In the gut, fluid shear stress induces for instance cell organization into 3D crypt and villi-like structure (Figure 1G). It has been shown thanks to a kidney model that fluid shear stress not only results in a morphological polarization triggering actin cytoskeleton reorganization, but also led to the expression of differentiated epithelial cell functions (Jang & Suh, 2010), (Theobald et al., 2018). Therefore, it appears capital to reproduce mechanical forces in 3D existing models. One of the main advantages that offer organ-on-chip is that tissue grown on a porous membrane can experience multiple physical stresses at the same time (Bhatia & Ingber, 2014). Furthermore, microorganisms have been also reported to sense physical cues such as the shear stress, reinforcing the needs to recapitulate physical forces to study host pathogen interactions. (DOI: 10.1016/j.cellsig.2020.109867 )

#### ***Stretching***

Organ-on-chip containing a porous polymer-based membrane can be easily stretched, but so far only few devices provide this mechanical stimulation. In the Emulate chip (Figure 1A), stretching is generated thanks to vacuum applied in two chambers located on both sides of the channels and can be applied either cyclically or continuously with big range of frequency and intensity. A negative pressure will apply stretching forces on the tissue whereas increasing the pressure inside the chambers will compress it. As an example, a cyclic stretch of 10% at 0.15Hz on a gut-on-chip recapitulates the peristaltic motion found in the human colon (H. J. Kim & Ingber, 2013).

### ***Shear forces***

Microfluidic devices are particularly relevant platforms to expose cells to fluid shear forces. In fact, laminar flow circulating in organ-on-chip naturally produces such forces on engineered tissue. As an example, primary kidney epithelial cells loaded into the apical side of a two-channels device and exposed to a fluid shear stress of  $0.2 \text{ dyne.cm}^{-2}$ , mimic the actual forces found in living kidney tubules and enhanced polarization and function of the cells (Jang et al., 2013).

### ***Electric field***

Some models include electrodes around the channels allowing the application of an electrical field within the tissue. The electrical stimulation of contractile cells helps reproducing electrophysiological and mechanical features of cardiac muscles for instance (Agarwal et al., 2013).

### ***Biochemical tension***

Organ-on-chip are suitable platforms to simulate and engineer a relevant chemical microenvironment. Gradients of nutrients or oxygen can be often observed in multiple human tissues but pathological tissues usually exhibit chemotaxis as well, in tumor cells for instance (Huh et al., 2012), or in heart ischemia characterized by a strong lack of oxygen inside the tissue (Khanal et al., 2011). Tumor-stromal interaction was engineered into a microfluidic chip designed to generate a biochemical tension, physiological chemotactic gradients influencing cancer cells migrations (Torisawa et al., 2010). In addition, the gut microbiota should differ upon aerobic or anaerobic conditions. Therefore, several groups could recreate an oxygen gradient in gut-on-chips supporting anaerobes co-culture (Figure 1C, 1F, (R. Kim et al., 2019), (Shah et al., 2016)).

### ***2-3-Representing specific 3D topology, interface and cellular richness***

One of the main challenges of tissue engineering is to recapitulate the cellular and molecular composition of the organ. In that respect stem cells such as induced-pluripotent stem cells (iPS) or even embryonic stem cells has been used (Wan et al., 2011). Many studies related to organ-on-chip have considered cell patterning as an efficient method to create precise cell environment with specific arrangement inside the chip. Surface treatments or even 3D printing guarantee precise patterns or complex hydrogels scaffold which can be eventually grown with cells. A recent study uses cell patterning to develop a hydrogel-based mini-gut model shaped with microcavities that recapitulate the 3D geometry and cell diversity (by loading organoids) of the native crypt and villi structure of the human gut (Figure 1B, (Nikolaev et al., 2020)). This

Similarly, a collagen-based array of crypt-like invaginations could closely mimic the human intestinal epithelium monolayer with a gradient of proliferative to differentiated cells along crypts (Figure 1C), (R. Kim et al., 2019) (Wang et al., 2018). These two devices enable the maintenance and spatial organization of both stem and differentiated cells and hence represent good models to investigate gut regeneration.

A tissue exhibits specific interface and needs a precise cell organization to be fully functional and to interact with the rest of the body. Human body displays several of them in many organs: lungs, intestine, vessels, bladder, skin, vagina are the most important. Pathogens enter the body through the airflow or through contaminated food and fluids. Organ-on-chips having two adjacent channels separated by a porous membrane offer many options in terms of cellular seeding and medium type (air, liquid or gas). Indeed, different kind of cells can be cultured on both sides of the porous membrane with their respective physiological media, allowing hence exchanges between several types of tissues/cells (Figure 1, graphical abstract). Lung is an appropriate example of tissue interface presenting an air-liquid boundary with relevant alveolar barrier that has been recapitulated in several chips designed by Emulate, Alevolix or Epithelix (Figure 1A, 1E, DOI: 10.1038/s41598-018-32523-x ). The intestinal lumen is another example of tissue liquid-liquid interface (Figure 1A, B, C). In addition, several devices allow at recapitulating several tissue-tissue interface as exemplified in Figure 1A, D, F.

### **3-Organ-on-chip to study host-pathogen/commensal interactions**

Respiratory tissues, gastrointestinal tract, skin or even mucosal surfaces are tissues that most likely encounter pathogens since they are in direct contact with outside the body. They are precisely considered as the first barriers against pathogens. Respiratory pathogens have to cross several barriers (mediators and bactericidal peptides) before reaching and infecting their target cells (Grubor et al., 2006). Concerning enteric pathogens, they experience strong pH changes and oxygen gradient tensions once ingested. Plasma membrane is the precise location where host and pathogens meet and is composed of several mechano-sensitive structures: caveola, lipid rafts, cytoskeleton components (actin fibers, tubulin transducing signals to mechanosensitive proteins). Indeed, Caco-2 cells grown in gut-on-chip or in Transwell device have a totally different profile of gene expression, reinforcing the impact of mechanical stimulation on tissue molecular organization (Kasendra et al., 2020). Given the great importance of mechanosensing in eukaryotic cells, it is not surprising that pathogens might experience such phenomenon as well.

#### ***3-1-Mechanical forces trigger pathogen virulence induction***

In this part we will briefly describe several studies reporting the induction of virulent genes by mechanical stimulation that were obtained with microfluidic devices (not truly organ-on-chip). Bacteria often experience mechanical forces when they encounter cell surfaces. They can be triggered by a fluid shear stress generated by a flow or be stimulated by the host surface stiffness and topology. It has been shown that bacteria could actually feel such signals and induce the upregulation or downregulation of their virulence factors.

Enterohemorrhagic *Escherichia coli* (EHEC), a foodborne pathogen that can colonize the gastrointestinal tract, was capable to adapt the expression of its virulence genes depending on the mechanical constraints of its host (Alsharif et al., 2015). More precisely, the locus of enterocyte effacement containing a range of virulence factors, notably the transcriptional activator GrlA, was shown highly expressed and localized into the cytoplasm in response to physiological relevant shear stress (Sirisaengtaksin et al., 2020).

Type IV pili are long and thin dynamic filaments based at the surface of bacteria and often involved in pathogenic processes. These surface appendages can polymerize and depolymerize from pilin molecules in a short period of time (Craig et al., 2019). Biomechanical considerations have revealed that type IV pili had an influence on how bacteria interact with host surfaces but also how they can be modulated by external mechanical stresses. They are actually considered by the literature as mechanosensitive structure. The case of *Neisseria meningitides*, a pathogenic bacterium infecting the human nasopharynx is worth noting. When this pathogen reaches the blood-brain barrier and integrates the bloodstream, it can cause severe inflammation named meningitis. Several studies have shown that despite the strong mechanical forces generated by the blood flow, *Neisseria* succeed to reach the bloodstream, bind to the epithelial vessel structure and aggregate, confirming their striking abilities of adaptation (Dos Santos Souza et al., 2020). *N. meningitides* aggregate in microcolonies, very stable and flow-resistant which induced host cell surface reorganization and protrusion formation enhancing the adhesion (DOI: 10.1016/j.ceb.2016.01.005 ). Interestingly, *in vivo* data showed that nesserial adhesion was mainly observed in capillaries, having heterogeneous flow, compared to normal-sized vessels and this would favor the initial adhesion of the bacterium (Mairey et al., 2006).

Biofilms are communities of bacteria held together with a specific extracellular matrix. The biofilm is created when a bacterium adheres to a surface. Interestingly, pathogens developing within a biofilm are phenotypically different from their free-living counterparts. It has been demonstrated that intracellular levels of cyclic-di-GMP participate to the initiation of *Pseudomonas aeruginosa* biofilms, such signals inducing a change in genes expression

(Rodesney et al., 2017). In this context, the role of mechanical forces such as shear forces would trigger bacterial behavior, enhancing the surface adhesion of *P. aeruginosa*, activating, via PilY1 and type IV pili, the cyclic-di-GMP pathway decisive for biofilm development (Rodesney et al., 2017).

The direct role of mechanical inputs of tissues on pathogenic gene expression has not yet been clearly investigated using organ-on-chip, but will surely be addressed in the close future.

### ***3-2-Organ-on-chip to model host-microbiota crosstalk***

Microbial community is a major part of the cellular richness of some tissues, especially in the gut where the number of bacteria can reach  $\sim 10^{14}$  bacteria (Qin et al., 2010), (Turnbaugh et al., 2007), (Lozupone et al., 2012). The intestinal tract composition varies according to space and time. The interaction between microbiota and the organ participate to tissue homeostasis and function, but also turns out to be a strong protection against pathogens. Indeed, several protective mechanisms such as mucin secretion, or the triggering of bacteria sensors are controlled by certain entities of the microbiota (Ashida et al., 2012). Classical devices exhibit several drawbacks to co-cultivate eukaryotic cell monolayers with microorganisms. In fact, they fail to recapitulate the microbial diversity, the long-term microbiota interaction and the complexity of the underlying tissue structure. In this context, several studies have highlighted that organ-on-chip devices can sustain complex microbial communities growing in symbiosis with human intestinal cells, relevant to host-microorganism interaction investigations.

The lack of fluid-flow in classical devices was thought to be the main missing cue leading to bacterial overgrowth. However, peristaltic motion is another important parameter, since cyclic stretching (peristalsis-like deformations) greatly reduces bacterial overgrowth in a gut-on-chip already maintained under luminal flow (H. J. Kim et al., 2016). This suggests that modulation of mechanical deformation is decisive to prevent bacterial overgrowth (Bures et al., 2010), and that gut-on-chip are reliable models to co-culture commensal microorganisms over an extended period of time. As an example, viable bacterial microcolonies from a commercial probiotic formulation containing several strains (*Bifidobacterium breve*, *B. longum*, *B. infantis*, *Lactobacillus acidophilus*, *L. plantarum*, *L. paracasei*, *L. bulgaricus*, and *Streptococcus thermophiles*) have been successfully maintained for several days (H. J. Kim et al., 2016).

In addition, hypoxia gradient in gut-on-chip greatly increase the microbial diversity since it increase the growth of facultative and obligate anaerobic bacteria like *Bacteroides fragilis* (Jalili-Firoozinezhad et al., 2019) or *B. adolescentis* and *L. rhamnosus*, (R. Kim et al., 2019). In particular, co-cultivating in anaerobic conditions *L. rhamnosus* with human colon cells in

HuMiX device was able to modify the transcriptional, metabolic and immunological host responses as reported previously by in vivo studies (Shah et al., 2016).

Altogether, these studies demonstrate the huge benefit of using gut-on-chip under mechanical stimulation and oxygen gradient to provide a microbiota-intestine-on-chip relevant model.

### ***3-3-Pathogenic mechanisms deciphered by organ-on-chip devices***

Over the last few years, host-pathogen interactions have been increasingly investigated supported by organ-on-chip technology, some examples (non exhaustive) are described below.

#### **Digestive system infections**

The gut is often targeted by pathogens and thus several studies use gut-on-chip devices. Gut-on-chip (Figure 1A) was used to monitor enterohemorrhagic *E. coli* (EHEC) infection in the presence of human-derived microbiome metabolites. The study highlighted that some metabolites were associated to a greater injury following EHEC infection, by inducing flagellin expression in the bacterium genome (Tovaglieri et al., 2019). Human enteroviruses infections have also been explored with organ-on-chip. For instance, the enterovirus coxsackievirus B1 (CVB1) injected to the lumen of gut-on-chip was able to enter the epithelium, and infectious virions and inflammatory cytokines were released in the apical exit fluid (Villeneuve et al., 2017). Another example is *Shigella*, a human restricted bacterium leading to severe intestinal damages. *S. flexneri* apical invasion was 10 000 fold more efficient in a gut-on-chip than in Transwell classical models, recapitulating for the first time the high infectivity of this bacterium as seen in human patients, highlighting once again the benefits of organ-on-chip devices in modeling human infection (Grassart et al., 2019). Furthermore, the study showed that the 3D architecture of the colon recapitulated in the gut-on-chip provided crypt-like structures where *S. flexneri* were highly enriched creating new infection outbreak points (Figure 1H). In addition, *Shigella* invasion and propagation through the epithelial monolayer were greatly enhanced by cyclic stretching and continuous flow. Altogether, this study reveals that *Shigella* takes advantage of the intestinal microarchitecture and mechanical forces found in the human colon to invade with great effectiveness the intestinal barrier (Grassart et al., 2019). How the physiological stretch enhances *Shigella* invasion is not yet known. Mechanical stimulation may induce the reorganization of the actin cytoskeleton and/or plasma membrane cholesterol, two known host cell factors needed for bacterial uptake, these forces may also directly induce virulent genes of *Shigella* (Graphical abstract).

Liver can be also recapitulated in organ-on-chip and hence allowing hepatitis B virus (HBV) infection. Ortega-Prieto et al designed a 3D microfluidic system made of collagen-coated

polystyrene scaffold continuously perfused with oxygenated medium, allowing the culture of primary human hepatocytes in fully polarized manner. This platform could recapitulate the innate immune and cytokine responses mimicking the HBV-infected patient's response (Ortega-Prieto et al., 2018). Future investigations using this platform could help identifying pathways implicated in the host-HBV interaction.

### ***Lung infections***

The lung is also a target organ for pathogens and in particular bacteria. A sophisticated organotypic bronchiole model was engineered to provide multiple interfaces embedded in a 3D matrix of collagen and pulmonary fibroblasts. The center lumen composed of primary human bronchial epithelial cells, filled with air to form an air-liquid interface, is flanked with two vascular compartments filled with primary human lung microvascular cells, that can be additionally loaded with polymorphonuclear leukocyte (PMN). This complex model enabled the authors to perform volatile administrations of *Aspergillus fumigatus* and/or *Pseudomonas aeruginosa* and observed the efficient (co-)infection by these pathogens, the inflammatory response of host tissues and PMN extravasation <https://www.nature.com/articles/s41467-017-01985-4>. Another work describes the use of an alveolus-on-chip to perform co-infection by *Staphylococcus aureus* and influenza virus. Thanks to organ-on-chip device the authors could mimic the superinfection seen in severe case of pneumonia (<https://iopscience.iop.org/article/10.1088/1758-5090/ab7073>). Pulmonary surfactant was proposed to play a host-protective role during the early interaction with *Mycobacterium tuberculosis*, however no animal model could accurately demonstrate this due the lethality of surfactant-deficient animals. Thanks to a lung-on-chip device, with air-liquid interface, having on the top channel alveolar epithelial cells and macrophages and at the bottom channel, endothelial cells, the authors investigated early steps of *M. tuberculosis* invasion and revealed the protective role of surfactants in tuberculosis (<https://elifesciences.org/articles/59961>). Another study used a microfluidic lung-on-chip with a constant fluid-flow to explore the role of bacteriophage-host interactions in mucosal immunity. Thanks to this mucus producing lung interface, the authors uncovered the sub-diffusive motion of T4 bacteriophage that was determinant to efficiently reduce *E coli* colonization (Barr et al., 2015).

Virus can also infect respiratory tissues, several studies assessed their invasion with organ-on-chip. The example of influenza virus is relevant since engineered lung-on-chip offers a reliable model of respiratory tissues including alveolus, small airway and alveolar-capillary for viral infection and drug development (Gkatzis et al., 2018). Indeed, a study employed such platforms

(Figure 1A) culturing primary human lung airway epithelial cells and inoculating different influenza virus strains (H1N1, H3N2, H5N1) (Si et al., 2019). The model reproduces faithfully the airborne influenza infection demonstrating high virulence accompanied by a barrier disruption and a loss of overlying cilia. Thus, the virulence of influenza virus and its possible mutations under antiviral drugs, through patient-patient transmission, could be also recapitulated using a lung-on-chip. Finally, the great adaptability of organ-on-chip devices allows investigations of emerging viral infections such as SARS-CoV-2 virus well illustrates. Several articles described the use of organ-on-chip to investigate SARS-CoV-2 infection. One study used the MucilAir platform (Epithelix) to observe the virus effect on primary human bronchial epithelial cells upon air-liquid interface. The authors observed a profound damage of the motile cilia function resulting in impaired mucociliary clearance (Robinot et al., 2020). Another study, using alveolar-on-chip reproduces key features of alveolar-capillary barrier by co-cultivating human alveolar epithelium, microvascular endothelium, and circulating immune cells under fluid-flow. The authors reported that high dose of SARS-CoV-2 leads to an efficient viral production in epithelial cells in contrast to endothelial cells, triggers immune cells recruitment, endothelium detachment, and inflammatory cytokines storm. <https://onlinelibrary.wiley.com/doi/epdf/10.1002/advs.202002928>. A similar alveolar-capillary barrier model was used with low dose of SARS-CoV-2, and revealed that even low viral replication still lead to an inflammatory response, and damage in endothelial layer integrity (DOI: 10.1101/2020.08.10.243220). While these two works have some discrepancy, they both highlight the differential response of epithelial and endothelial cells.

### ***Blood and bladder infections***

A simple microvessel-on-chip using OrganoPlate (Mimetas) with bidirectional flow was used to address hemorrhagic shock syndrome with Ebola <https://doi.org/10.1016/j.isci.2019.100765> and Lassa <https://doi.org/10.1002/bit.27636> viruses. The luminal infusion of Lassa or Ebola virus caused a dramatic effect on vascular permeability.

The urinary tract is another tissue often targeted by pathogens, Uropathogenic *E. coli* (UPEC) is one such example. Using a bladder-chip (same design than Figure 1A) having bladder epithelial cells and microvascular endothelial cells under flow in urine and nutritive media, and recapitulating bladder filling and voiding cycles, they investigate the formation of intracellular bacterial communities (IBCs) by time-lapse microscopy. Their results suggest that, IBCs may play a role in reseeding sites of infection for a long period after antibiotic treatment initiation DOI: 10.1101/2021

## **5-Conclusion**

Although recently developed, few studies employing organ-on-chip have demonstrated the high potential of this technology to understand infectious processes. By recapitulating the tissue physiology, with its precise physical cues and interface, organ-on-chip allow at addressing new questions or revisit old dogma in host-pathogen interactions that will for sure unravel key findings in the future. Beside, this method proposes a way to reduce animal experimentation required by the international community. Finally, the emergence of body-on-chip using human derived organoids, will be a breakthrough towards personalized medicine and drug testing.

## **Acknowledgement**

We are very grateful to Assu2000 and PTR232-2019 to fund our team and the fellowship of Thomas Feugas. We are particularly thankful to the Biomaterial and Microfluidic core facility (BMcf) of Institut Pasteur headed by Samy Gobaa and in particular to Heloise Mary for training us on organ-on-chip and for sharing her knowledge on this technology. We would like to thank Emulate Company and in particular Brett Clair and Stan King for providing us some images. We thank the PBI (Imagopole) platform of Institut Pasteur for microscope maintenance and technical help.

## **Conflict of interest**

We declare no conflict of interest

## **References**

- Agarwal, A., Goss, J. A., Cho, A., McCain, M. L., & Parker, K. K. (2013). Microfluidic heart on a chip for higher throughput pharmacological studies. *Lab on a Chip*, 13(18), 3599. <https://doi.org/10.1039/c3lc50350j>
- Alsharif, G., Ahmad, S., Islam, M. S., Shah, R., Busby, S. J., & Krachler, A. M. (2015). Host attachment and fluid shear are integrated into a mechanical signal regulating virulence in *Escherichia coli* O157:H7. *Proceedings of the National Academy of Sciences of the United States of America*, 112(17), 5503–5508. <https://doi.org/10.1073/pnas.1422986112>
- Ashida, H., Ogawa, M., Kim, M., Mimuro, H., & Sasakawa, C. (2012). Bacteria and host interactions in the gut epithelial barrier. *Nature Chemical Biology*, 8(1), 36–45. <https://doi.org/10.1038/nchembio.741>

Barr, J. J., Auro, R., Sam-Soon, N., Kassegne, S., Peters, G., Bonilla, N., Hatay, M., Mourtada, S., Bailey, B., Youle, M., Felts, B., Baljon, A., Nulton, J., Salamon, P., & Rohwer, F. (2015). Subdiffusive motion of bacteriophage in mucosal surfaces increases the frequency of bacterial encounters. *Proceedings of the National Academy of Sciences*, 112(44), 13675–13680. <https://doi.org/10.1073/pnas.1508355112>

Barrila, J., Radtke, A. L., Crabbé, A., Sarker, S. F., Herbst-Kralovetz, M. M., Ott, C. M., & Nickerson, C. A. (2010). Organotypic 3D cell culture models : Using the rotating wall vessel to study host-pathogen interactions. *Nature Reviews. Microbiology*, 8(11), 791–801. <https://doi.org/10.1038/nrmicro2423>

Bhatia, S. N., & Ingber, D. E. (2014). Microfluidic organs-on-chips. *Nature Biotechnology*, 32(8), 760–772. <https://doi.org/10.1038/nbt.2989>

Bures, J., Cyrany, J., Kohoutova, D., Förstl, M., Rejchrt, S., Kvetina, J., Vorisek, V., & Kopacova, M. (2010). Small intestinal bacterial overgrowth syndrome. *World Journal of Gastroenterology*, 16(24), 2978–2990. <https://doi.org/10.3748/wjg.v16.i24.2978>

Craig, L., Forest, K. T., & Maier, B. (2019). Type IV pili : Dynamics, biophysics and functional consequences. *Nature Reviews. Microbiology*, 17(7), 429–440. <https://doi.org/10.1038/s41579-019-0195-4>

Ding, C., Chen, X., Kang, Q., & Yan, X. (2020). Biomedical Application of Functional Materials in Organ-on-a-Chip. *Frontiers in Bioengineering and Biotechnology*, 8, 823. <https://doi.org/10.3389/fbioe.2020.00823>

Dos Santos Souza, I., Maïssa, N., Ziveri, J., Morand, P. C., Coureuil, M., Nassif, X., & Bourdoulous, S. (2020). Meningococcal disease : A paradigm of type-IV pilus dependent pathogenesis. *Cellular Microbiology*, 22(4), e13185. <https://doi.org/10.1111/cmi.13185>

Esch, M. B., Mahler, G. J., Stokol, T., & Shuler, M. L. (2014). Body-on-a-chip simulation with gastrointestinal tract and liver tissues suggests that ingested nanoparticles have the potential to cause liver injury. *Lab on a Chip*, 14(16), 3081–3092. <https://doi.org/10.1039/c4lc00371c>

Gkatzis, K., Taghizadeh, S., Huh, D., Stainier, D. Y. R., & Bellusci, S. (2018). Use of three-dimensional organoids and lung-on-a-chip methods to study lung development, regeneration and disease. *The European Respiratory Journal*, 52(5). <https://doi.org/10.1183/13993003.00876-2018>

Grassart, A., Malardé, V., Gobaa, S., Sartori-Rupp, A., Kerns, J., Karalis, K., Marteyn, B., Sansonetti, P., & Sauvonnnet, N. (2019). Bioengineered Human Organ-on-Chip Reveals Intestinal

Microenvironment and Mechanical Forces Impacting Shigella Infection. *Cell Host & Microbe*, 26(3), 435-444.e4. <https://doi.org/10.1016/j.chom.2019.08.007>

Grubor, B., Meyerholz, D. K., & Ackermann, M. R. (2006). Collectins and cationic antimicrobial peptides of the respiratory epithelia. *Veterinary Pathology*, 43(5), 595-612. <https://doi.org/10.1354/vp.43-5-595>

Huh, D., Torisawa, Y., Hamilton, G. A., Kim, H. J., & Ingber, D. E. (2012). Microengineered physiological biomimicry: Organs-on-Chips. *Lab on a Chip*, 12(12), 2156. <https://doi.org/10.1039/c2lc40089h>

Jalili-Firoozinezhad, S., Gazzaniga, F. S., Calamari, E. L., Camacho, D. M., Fadel, C. W., Bein, A., Swenor, B., Nestor, B., Cronce, M. J., Tovaglieri, A., Levy, O., Gregory, K. E., Breault, D. T., Cabral, J. M. S., Kasper, D. L., Novak, R., & Ingber, D. E. (2019). A complex human gut microbiome cultured in an anaerobic intestine-on-a-chip. *Nature Biomedical Engineering*, 3(7), 520-531. <https://doi.org/10.1038/s41551-019-0397-0>

Jang, K.-J., Mehr, A. P., Hamilton, G. A., McPartlin, L. A., Chung, S., Suh, K.-Y., & Ingber, D. E. (2013). Human kidney proximal tubule-on-a-chip for drug transport and nephrotoxicity assessment. *Integrative Biology: Quantitative Biosciences from Nano to Macro*, 5(9), 1119-1129. <https://doi.org/10.1039/c3ib40049b>

Jang, K.-J., & Suh, K.-Y. (2010). A multi-layer microfluidic device for efficient culture and analysis of renal tubular cells. *Lab on a Chip*, 10(1), 36-42. <https://doi.org/10.1039/b907515a>

Kasendra, M., Luc, R., Yin, J., Manatakis, D. V., Kulkarni, G., Lucchesi, C., Sliz, J., Apostolou, A., Sunuwar, L., Obrigewitch, J., Jang, K.-J., Hamilton, G. A., Donowitz, M., & Karalis, K. (2020). Duodenum Intestine-Chip for preclinical drug assessment in a human relevant model. *ELife*, 9. <https://doi.org/10.7554/eLife.50135>

Khanal, G., Chung, K., Solis-Wever, X., Johnson, B., & Pappas, D. (2011). Ischemia/reperfusion injury of primary porcine cardiomyocytes in a low-shear microfluidic culture and analysis device. *The Analyst*, 136(17), 3519. <https://doi.org/10.1039/c0an00845a>

Kim, H. J., & Ingber, D. E. (2013). Gut-on-a-Chip microenvironment induces human intestinal cells to undergo villus differentiation. *Integrative Biology: Quantitative Biosciences from Nano to Macro*, 5(9), 1130-1140. <https://doi.org/10.1039/c3ib40126j>

Kim, H. J., Li, H., Collins, J. J., & Ingber, D. E. (2016). Contributions of microbiome and mechanical deformation to intestinal bacterial overgrowth and inflammation in a human gut-on-

a-chip. *Proceedings of the National Academy of Sciences*, 113(1), E7–E15. <https://doi.org/10.1073/pnas.1522193112>

Kim, R., Attayek, P. J., Wang, Y., Furtado, K. L., Tamayo, R., Sims, C. E., & Allbritton, N. L. (2019). An in vitro intestinal platform with a self-sustaining oxygen gradient to study the human gut/microbiome interface. *Biofabrication*, 12(1), 015006. <https://doi.org/10.1088/1758-5090/ab446e>

Lozupone, C. A., Stombaugh, J. I., Gordon, J. I., Jansson, J. K., & Knight, R. (2012). Diversity, stability and resilience of the human gut microbiota. *Nature*, 489(7415), 220–230. <https://doi.org/10.1038/nature11550>

Mairey, E., Genovesio, A., Donnadieu, E., Bernard, C., Jaubert, F., Pinard, E., Seylaz, J., Olivo-Marin, J.-C., Nassif, X., & Duménil, G. (2006). Cerebral microcirculation shear stress levels determine *Neisseria meningitidis* attachment sites along the blood-brain barrier. *The Journal of Experimental Medicine*, 203(8), 1939–1950. <https://doi.org/10.1084/jem.20060482>

Nikolaev, M., Mitrofanova, O., Broguiere, N., Geraldo, S., Dutta, D., Tabata, Y., Elci, B., Brandenberg, N., Kolotuev, I., Gjorevski, N., Clevers, H., & Lutolf, M. P. (2020). Homeostatic mini-intestines through scaffold-guided organoid morphogenesis. *Nature*, 585(7826), 574–578. <https://doi.org/10.1038/s41586-020-2724-8>

Ortega-Prieto, A. M., Skelton, J. K., Wai, S. N., Large, E., Lussignol, M., Vizcay-Barrena, G., Hughes, D., Fleck, R. A., Thursz, M., Catanese, M. T., & Dorner, M. (2018). 3D microfluidic liver cultures as a physiological preclinical tool for hepatitis B virus infection. *Nature Communications*, 9(1), 682. <https://doi.org/10.1038/s41467-018-02969-8>

Qin, J., Li, R., Raes, J., Arumugam, M., Burgdorf, K. S., Manichanh, C., Nielsen, T., Pons, N., Levenez, F., Yamada, T., Mende, D. R., Li, J., Xu, J., Li, S., Li, D., Cao, J., Wang, B., Liang, H., Zheng, H., ... Wang, J. (2010). A human gut microbial gene catalogue established by metagenomic sequencing. *Nature*, 464(7285), 59–65. <https://doi.org/10.1038/nature08821>

Robinot, R., Hubert, M., de Melo, G. D., Lazarini, F., Bruel, T., Smith, N., Levallois, S., Larrous, F., Fernandes, J., Gellenoncourt, S., Rigaud, S., Gorgette, O., Thouvenot, C., Trébeau, C., Mallet, A., Duménil, G., Gobaa, S., Etournay, R., Lledo, P.-M., ... Chakrabarti, L. A. (2020). SARS-CoV-2 infection damages airway motile cilia and impairs mucociliary clearance. *bioRxiv*, 2020.10.06.328369. <https://doi.org/10.1101/2020.10.06.328369>

Rodesney, C. A., Roman, B., Dhamani, N., Cooley, B. J., Katira, P., Touhami, A., & Gordon, V. D. (2017). Mechanosensing of shear by *Pseudomonas aeruginosa* leads to increased levels of the cyclic-di-GMP signal initiating biofilm development. *Proceedings of the National Academy of*

Sciences of the United States of America, 114(23), 5906–5911. <https://doi.org/10.1073/pnas.1703255114>

Rosales, A. M., Vega, S. L., DelRio, F. W., Burdick, J. A., & Anseth, K. S. (2017). Hydrogels with Reversible Mechanics to Probe Dynamic Cell Microenvironments. *Angewandte Chemie (International Ed. in English)*, 56(40), 12132–12136. <https://doi.org/10.1002/anie.201705684>

Seghir, R., & Arscott, S. (2015). Extended PDMS stiffness range for flexible systems. *Sensors and Actuators A: Physical*, 230, 33–39. <https://doi.org/10.1016/j.sna.2015.04.011>

Seliktar, D. (2012). Designing cell-compatible hydrogels for biomedical applications. *Science (New York, N.Y.)*, 336(6085), 1124–1128. <https://doi.org/10.1126/science.1214804>

Shah, P., Fritz, J. V., Glaab, E., Desai, M. S., Greenhalgh, K., Frachet, A., Niegowska, M., Estes, M., Jäger, C., Seguin-Devaux, C., Zenhausern, F., & Wilmes, P. (2016). A microfluidics-based in vitro model of the gastrointestinal human-microbe interface. *Nature Communications*, 7, 11535. <https://doi.org/10.1038/ncomms11535>

Si, L., Bai, H., Rodas, M., Cao, W., Oh, C. Y., Jiang, A., Moller, R., Hoagland, D., Oishi, K., Horiuchi, S., Uhl, S., Blanco-Melo, D., Albrecht, R. A., Liu, W.-C., Jordan, T., Nilsson-Payant, B. E., Logue, J., Haupt, R., McGrath, M., ... Ingber, D. E. (2020). Human organ chip-enabled pipeline to rapidly repurpose therapeutics during viral pandemics [Preprint]. *Microbiology*. <https://doi.org/10.1101/2020.04.13.039917>

Si, L., Prantil-Baun, R., Benam, K. H., Bai, H., Rodas, M., Burt, M., & Ingber, D. E. (2019). Discovery of influenza drug resistance mutations and host therapeutic targets using a human airway chip [Preprint]. *Bioengineering*. <https://doi.org/10.1101/685552>

Simian, M., & Bissell, M. J. (2017). Organoids: A historical perspective of thinking in three dimensions. *Journal of Cell Biology*, 216(1), 31–40. <https://doi.org/10.1083/jcb.201610056>

Sirisaengtaksin, N., Odem, M. A., Bosserman, R. E., Flores, E. M., & Krachler, A. M. (2020). The E. coli transcription factor GrlA is regulated by subcellular compartmentalization and activated in response to mechanical stimuli. *Proceedings of the National Academy of Sciences of the United States of America*, 117(17), 9519–9528. <https://doi.org/10.1073/pnas.1917500117>

Sontheimer-Phelps, A., Hassell, B. A., & Ingber, D. E. (2019). Modelling cancer in microfluidic human organs-on-chips. *Nature Reviews Cancer*, 19(2), 65–81. <https://doi.org/10.1038/s41568-018-0104-6>

Theobald, J., Ghanem, A., Wallisch, P., Banaeiyan, A. A., Andrade-Navarro, M. A., Taškova, K., Haltmeier, M., Kurtz, A., Becker, H., Reuter, S., Mrowka, R., Cheng, X., & Wölfl, S. (2018).

Liver-Kidney-on-Chip To Study Toxicity of Drug Metabolites. *ACS Biomaterials Science & Engineering*, 4(1), 78–89. <https://doi.org/10.1021/acsbiomaterials.7b00417>

Torisawa, Y.-S., Mosadegh, B., Bersano-Begey, T., Steele, J. M., Luker, K. E., Luker, G. D., & Takayama, S. (2010). Microfluidic platform for chemotaxis in gradients formed by CXCL12 source-sink cells. *Integrative Biology: Quantitative Biosciences from Nano to Macro*, 2(11–12), 680–686. <https://doi.org/10.1039/c0ib00041h>

Tovaglieri, A., Sontheimer-Phelps, A., Geirnaert, A., Prantil-Baun, R., Camacho, D. M., Chou, D. B., Jalili-Firoozinezhad, S., de Wouters, T., Kasendra, M., Super, M., Cartwright, M. J., Richmond, C. A., Breault, D. T., Lacroix, C., & Ingber, D. E. (2019). Species-specific enhancement of enterohemorrhagic *E. coli* pathogenesis mediated by microbiome metabolites. *Microbiome*, 7(1), 43. <https://doi.org/10.1186/s40168-019-0650-5>

Turnbaugh, P. J., Ley, R. E., Hamady, M., Fraser-Liggett, C. M., Knight, R., & Gordon, J. I. (2007). The Human Microbiome Project. *Nature*, 449(7164), 804–810. <https://doi.org/10.1038/nature06244>

Villeneuve, R., Wales, S. Q., Hamkins-Indik, T., Papafragkou, E., Weaver, J. C., Ferrante, T. C., Bahinski, A., Elkins, C. A., Kulka, M., & Ingber, D. E. (2017). Human Gut-On-A-Chip Supports Polarized Infection of Coxsackie B1 Virus In Vitro. *PloS One*, 12(2), e0169412. <https://doi.org/10.1371/journal.pone.0169412>

Wan, C., Chung, S., & Kamm, R. D. (2011). Differentiation of embryonic stem cells into cardiomyocytes in a compliant microfluidic system. *Annals of Biomedical Engineering*, 39(6), 1840–1847. <https://doi.org/10.1007/s10439-011-0275-8>

Wang, Y., Kim, R., Gunasekara, D. B., Reed, M. I., DiSalvo, M., Nguyen, D. L., Bultman, S. J., Sims, C. E., Magness, S. T., & Allbritton, N. L. (2018). Formation of Human Colonic Crypt Array by Application of Chemical Gradients Across a Shaped Epithelial Monolayer. *Cellular and Molecular Gastroenterology and Hepatology*, 5(2), 113–130. <https://doi.org/10.1016/j.jcmgh.2017.10.007>

Weeber, F., Ooft, S. N., Dijkstra, K. K., & Voest, E. E. (2017). Tumor Organoids as a Pre-clinical Cancer Model for Drug Discovery. *Cell Chemical Biology*, 24(9), 1092–1100. <https://doi.org/10.1016/j.chembiol.2017.06.012>

Zhang, B., Lai, B. F. L., Xie, R., Davenport Huyer, L., Montgomery, M., & Radisic, M. (2018). Microfabrication of AngioChip, a biodegradable polymer scaffold with microfluidic vasculature. *Nature Protocols*, 13(8), 1793–1813. <https://doi.org/10.1038/s41596-018-0015-8>

Zhang, B., & Radisic, M. (2017). Organ-on-a-chip devices advance to market. *Lab on a Chip*, 17(14), 2395–2420. <https://doi.org/10.1039/c6lc01554a>


Zhao, H., Xu, K., Zhu, P., Wang, C., & Chi, Q. (2019). Smart hydrogels with high tunability of stiffness as a biomimetic cell carrier. *Cell Biology International*, 43(2), 84–97. <https://doi.org/10.1002/cbin.11091>

## Figure legends

Figure 1. Organ-on-chip models with their specific features: number of channel, material used, physical force recapitulated, possible interface and 3D/2D topology. (A) Commercial device from Emulate company, Intestine/Lung/Liver/Kidney-chip available. (B) Mini gut fabricated by micropatterning (Nikolaev et al., 2020). (C) Engineered crypts surrogates recapitulating the 3D architecture and oxygen gradient of the intestinal crypt (R. Kim et al., 2019). (D) AngioChip modelling a vessel on chip (Zhang et al., 2018). (E) Epithelix device recapitulating a airway-on-chip with specific air-liquid interface. (F) HuMix model (Shah et al., 2016). (G) 3D topology of a gut-on-chip with a crypt and villi structure from Emulate Intestine-Chip. (H) Cross section of gut-on-chip infected by the wild type *Shigella flexneri* strain, enriched in the intestinal crypts (Thomas Feaugas personal communication).


Figure 1

**A** Intestine-Chip (Emulate)




- 2 channels
- PDMS-based
- Stretch, shear forces, Oxygen tension
- Air-Liquid/Liquid-Liquid interface
- 3D topology

**B** Mini gut




- 1 channel
- Hydrogel based
- Shear forces
- Liquid-Liquid interface
- 3D topology

**C** Engineered crypt




- 1 channel
- Hydrogel based
- Oxygen tension
- Liquid-Liquid interface
- 3D topology

**D** AngioChip




- 1 channel
- PDMS/Hydrogel based
- Shear forces
- Liquid-Liquid interface
- 3D topology

**E** Airway-on-chip (Epithelix)



- 1 channel
- PTFE-based
- No physical forces
- Air-Liquid interface
- 2D topology

**F** HuMix model



- 3 channels
- Polycarbonate-based
- Shear forces
- Liquid-Liquid interface
- 2D topology

